

## **APHRS Advisory Notice 6<sup>th</sup> December 2020**

### **Device:**

EMBLEM™ Subcutaneous Implantable Cardioverter Defibrillators (S-ICDs) (Models A209 and A219) all Implanted from May 2015 to December 2017.

### **Description:**

Potential for electrical overstress and non-delivery of high voltage shock. Over time, due to variations in header assembly, a very small pathway may develop that allows moisture ingress, enabling a shorting condition to occur during delivery of high voltage therapy.

### **Presentation:**

Patient may present with failure of delivery high voltage therapies due to electrical short and subsequent necessity of device replacement post high voltage therapies which have shorted. Most common presentation is early device replacement and no clinical event. An occurrence of electrical overstress malfunction can be identified by the inability to perform a device interrogation (in-clinic or remotely via LATITUDE) or by device-based errors/alerts.

### **Rate of occurrence:**

The current projected occurrence rate for this advisory is 0.3% at 5 years, and the most common clinical outcome is early device replacement. To date there have been no serious sequelae, the potential exists for life-threatening harm due to an inability to provide defibrillation therapy. It is estimated the probability of the hypothetical worst-case harm associated with loss of therapy resulting in death is 0.09% at 5 years. Six confirmed events resulting in early replacement have occurred, four were reported as inability to interrogate, one displayed prolonged charge time alerts, and one exhibited premature battery depletion.

### **Recommendation:**

Increase device follow-up to 3 months in person or enrol appropriate patients into home monitoring with 3 monthly checks. Educate patients regarding beeping tones, if their device makes a beeping tone noise the patient must seek medical advice and device interrogation as soon as possible.

Patients at extremely high risk (i.e frequent appropriate device treatment events) and the patient cannot undergo increased monitoring as above the patient could be considered for device replacement at the physician's discretion, acknowledging the risks of complication associated with replacement are not insignificant.